



अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR

Address: Plot No. 2, Sector-20, MIHAN, Nagpur- 441124

Email - hospitalstore@aiimsnagpur.edu.in

GSTIN:27AAHAA5072G2ZS

File No. ADMIN-18.0/210/2025-O/O AO(ADMIN)-AIIMS-NAGPUR/3409/NIQ/39

Date: 28/08/2026

NOTICE INVITING QUOTATIONS

The Executive Director, All India Institute of Medical Sciences, Nagpur invites quotations from the reputed supplier on their letter head for supply of "Process Challenge Device (PCD)" for CSSD Department of AIIMS Nagpur mentioning the rate in following format:

Sr. No	Item Description	QTY	Unit Cost	Total Cost	GST %	Total Cost with GST
1	PCD Steam (BMS) Compact-PCD Compactable to EN 285-Green Hollow Load Test	06 Nos				
2.	PCD-Bowie-dick Compact-PCD-Blue EU-Version, Stimulating 7kg Cotton pack, Compatible to EN-285, 11140-4 & Hollow Load Test.	06 Nos.				
3.	PCD Biological Steam B-PM-RCPCD Compactable to EN 11138-1+3	06 Nos.				
Technical Specification attached: (Annexure-A)						
Total----						

Note: - L1 will be decided item wise including GST

TERMS OF CONDITIONS:

1. All quotations to be submitted in the name of Executive Director, AIIMS Nagpur only. Quotations not addressed to Executive Director, AIIMS, Nagpur will not be opened and rejected summarily.
2. Copy of NIQ duly signed and stamped by the vendor to be submitted along with quotation by accepting all the terms and conditions of NIQ.
3. Rates of articles should be quoted per unit item without including GST.
4. Sample may be obtained from the vendor, if considered necessary.
5. Validity of quotation should be minimum 60 days from the date of quotation.
6. Taxes leviable should be quoted separately.
7. Delivery Period: - Within 30 days from the date of supply order.
8. Delivery should be FOR Destination i.e. AIIMS, Nagpur
9. Liquidated Damages charges: - In the event of the Seller's failure supply, Liquidated Damages charges will be deducted to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part their of a week, subject to the maximum value of the Liquidated Damages being not higher that 10% of the value of delayed stores.
10. Payment will be released on inspection of supplied material found satisfactory by the concerned department as regard to quality, quantity etc. Broken articles will have to be replaced before payment.
11. Bank details such as Account Number, IFSC Code etc. should be furnished so as to facilitate payment online, if any.
12. A declaration by vendor is required to be submitted along with quotation stating that "vendor is not debarred by Department of Commerce or Ministry / Department concerned". The date of declaration should not be before the date of NIQ and after the last date of submission of quotation.
13. Interested parties should submit their quotations within 07 days from the date of issue of the NIQ. The quotations should be sent in sealed envelope in "Central Medical Store, 1st Floor, IPD Building, AIIMS Nagpur", superscribing on envelop "QUOTATION FOR SUPPLY OF "Process Challenge Device (PCD)" FOR CSSD DEPARTMENT WITH NUMBER OF NIQ" within working hour of Institute. If last date is Sunday or holiday, next working day will be considered as last date of submission.
14. Rights to accept/reject any quotation rests with the AIIMS Nagpur.

विपिन कुमार/Vipin Kumar 27.08.2026
(Shri Vipin Kumar)
प्रशासनिक अधिकारी (पी & एमएम)/Administrative Officer (P & MM)
अ.म.आ.स. नागपुर/AIIMS, Nagpur
AIIMS, Nagpur

Annexure-A

Technical Specification of the PCD Steam Sterilization

1. Process Challenge Device (PCD)-Steam (BMS)

- Compact reusable Process Challenge Device (PCD) for routine steam sterilization monitoring.
- Compatible with pre-vacuum steam sterilizers.
- Shall be designed for Hollow Load Testing in accordance with EN 285.
- Shall be compatible with **Green Hollow Load Test** indicator strips.
- Reusable stainless steel or equivalent durable construction resistant to repeated sterilization cycles.
- Suitable for routine daily monitoring and qualification of steam sterilizers.
- CDSCO approved.

2. Process challenge Device (PCD)- Bowie-Dick Test

- Compact reusable Bowie-Dick Process Challenge Device.
- Blue EU Version.
- Shall simulate a **7kg cotton pack** challenge.
- Compatible with pre-vacuum steam sterilizers.
- Shall comply with **EN 285**.
- Shall comply with **EN ISO 11140-4**.
- **Compatible with commercially available Bowie-Dick indicator sheets specified by the manufacturer.**
- CDSCO Approved

3. Biological Process challenge Device (PCD)- Steam

- Reusable Biological Process Challenge Device (B-PM-RCPCD) for steam sterilization.
- **Compatible with Biological Indicators used for steam sterilization monitoring (Instant Bio Plush SCBI 134 degree).**
- Shall comply with **EN ISO 11138-1** and **EN ISO 11138-3**.
- Designed to provide a standardized challenge equivalent to difficult-to-sterilize loads.
- Suitable for routine monitoring, validation, and qualification of steam sterilizers.
- CDSCO Approved.