



अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR

Address: Plot No. 2, Sector-20, MIHAN, Nagpur- 441124

Email - hospitalstore@aiimsnagpur.edu.in

File No. ADMIN-18.0/122/2026-O/O AO(ADMIN)-AIIMS-NAGPUR /4029/NIQ/13

Date: 17/06/2026

NOTICE INVITING QUOTATIONS

The Executive Director, All India Institute of Medical Sciences, Nagpur invites quotations from the reputed supplier on their letter head for **Procurement of Biological Indicators for Department of CSSD** at AIIMS Nagpur mentioning the rate as given below-

| Sr. No | Details of Item | Qty. | Unit Cost | Total Cost | GST % | Total Cost with GST |
|--------|--|--|-----------|------------|-------|---------------------|
| 1 | Super Rapid Biological Indicator (Technical specification as per Annexure-I) | 14 Boxes (Pack Size- 100 Nos. per Box) | | | | |
| 2 | Rapid readout EO Bio-Ind Sterile (Technical specification as per Annexure-I) | 5 Boxes (Pack Size- 100 Nos. per Box) | | | | |

Note- L1 will be decided item wise including GST

TERMS OF CONDITIONS:

- All quotations to be submitted in the name of Executive Director, AIIMS Nagpur only. Quotations not addressed to Executive Director, AIIMS, Nagpur will not be opened and rejected summarily.
- Copy of NIQ duly signed and stamped by the vendor to be submitted along with quotation by accepting all the terms and conditions of NIQ.**
- Rates of articles should be quoted per unit item without including GST.
- Sample may be obtained from the vendor, if required by user department.**
- Please, submit the quotation along with relevant Brochure & catalogue mentioning the detailed specification.**
- Validity of quotation should be minimum 60 days from the date of quotation.**
- Taxes leviable should be quoted separately.
- Delivery Period: - Within 30 days from the date of supply order.
- Delivery should be FOR Destination i.e. AIIMS, Nagpur
- Liquidated Damages charges: - In the event of the Seller's failure supply, Liquidated Damages charges will be deducted to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part their of a week, subject to the maximum value of the Liquidated Damages being not higher that 10% of the value of delayed stores.
- Payment will be released on inspection of supplied material found satisfactory by the concerned department as regard to quality, quantity etc. Broken articles will have to be replaced before payment.
- Bank details such as Account Number, IFSC Code etc. should be furnished so as to facilitate payment online, if any.
- A declaration by vendor is required to be submitted on firm's letter head duly signed & stamped along with quotation stating that "vendor is not debarred by Department of Commerce or Ministry / Department concerned". The date of declaration should not be before the date of NIQ and after the last date of submission of quotation.**
- Interested parties should submit their quotations **within 07 days from the date of issue of the NIQ.** The quotations should be sent in sealed envelope in "Central Medical Store, 1st Floor, IPD Building, AIIMS Nagpur", superscribing on envelop "QUOTATION FOR PROCUREMENT OF BIOLOGICAL INDICATORS FOR DEPARTMENT OF CSSD WITH NUMBER OF NIQ" within working hour of Institute. If last date is Sunday or holiday, next working day will be considered as last date of submission.
- Rights to accept/reject any quotation rests with the AIIMS Nagpur.



Administrative Officer

AIIMS, Nagpur

शब्बीर शेख / Shabbir Sheikh
प्रशासनिक अधिकारी / Administrative Officer

ब.भा.आ.सं. नागपुर/AIIMS Nagpur

GENERAL SPECIFICATIONS:

- Should have FDA or ISO or EN ISO certification
- Should contain Quality Assurance certificate with mention of Spore population, D- value & Z value (preferably with each box).
- Self-life should be a minimum of 12–18 months. Batch-wise COA (Certificate of Analysis) must be provided.
- **If any reader/incubator is required for interpretation of BI results, it should be provided by the vendor.**
- **Demo Samples to be sent to check for quality of the product.**
- **Demonstration to be conducted in CSSD prior to technical approval.**
- **Fastest read out time will be given preference**
- **Read out time should be less than 12hr.**

BIOLOGICAL INDICATOR – STEAM STERILIZATION

1. Should be Self-Contained Biological Indicator (SCBI) for steam sterilization. **(Read-out time to be specified)**
2. Each ampule should contain *Geobacillus stearothermophilus* spores (population should be at least 10^6 CFU).
3. Biological indicator should comply with EN ISO 11138-1 and EN ISO 11138-3
4. Each ampule should have inbuilt Type 5 Chemical Indicator for instant release.
5. Type 5 indicator should comply with EN ISO 11140-2
6. Should be suitable for Pre-Vacuum Steam Cycles (132°C to 135°C).
7. Should be compatible for use in Compact Process Challenge Device (PCD) complying with EN ISO 11140-6.
8. Should be CDSCO approved.

BIOLOGICAL INDICATOR – ETHYLENE OXIDE (ETO) STERILIZATION

1. Should be Self-Contained Biological Indicator (SCBI) for EO sterilization. **(Read-out time to be specified)**
2. Each ampule should contain *Bacillus atrophaeus* spores (population should be at least 10^6 CFU).
3. Should comply with EN ISO 11138-2
4. Should be suitable for EO sterilization cycles at 36°C to 50°C .
5. Should be compatible with validated Process Challenge Device (PCD) complying with EN 1422
6. Should be CDSCO approved.

