


Specification:

BIOLOGICAL INDICATOR AUTOREADER 230VAC

- Instrument Incubator should be compatible for both Pre Vacuum and Gravity Steam Sterilizers
- Instrument should have a built-in web app to allow create documentation that can be printed or saved electronically.
- Equipment should show remaining minutes left of BI incubation.
- Heating block temperature reading should be displayed
- Positive BI Read-out time within 1 hr by an audible alarm.
- Auto reader, self calibrating, to read Rapid Attest BI, with ten incubation wells
- Compatible BI should be FDA cleared
- Equipment should be suitable for Biological Indicator Steam Sterilizer and Plasma Sterilization
- Equipment should have warranty of one year
- Vendor should provide 50 BI for steam sterilizer and 50 BI for plasma sterilizer free of cost compatible with the equipment.

Certification/ Validation requirements:

- Equipment should have IEC 61326-1 Electrical equipment for measurement, control and laboratory use -EMC requirements - Part 1: General requirements
- Equipment should comply with the EMsC requirements of the CE mark EMC Directive 2004/108/EC.


Dr. Meena Mishra
Professor & HOD
Department of Microbiology
AIIMS, Nagpur