



अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR  
Address: Plot No. 2, Sector-20, MIHAN, Nagpur- 441108  
Email: - [procurement@aiimsnagpur.edu.in](mailto:procurement@aiimsnagpur.edu.in), Ph.No.0710-3295591



F. No. AIIMS-NGP/PROC/HOSP/NIQ/07/23/74

Date. 20-07-2023

**NOTICE INVITING QUOTATIONS**

The Director, All India Institute of Medical Sciences, Nagpur invites quotations for the Oral & Maxillofacial Cross Drive Max Kit to Providing Oral & Maxillofacial Services in the Department of Dentistry, AIIMS Nagpur.

Details are given below:


Sr. No.	Items Name	Qty.	Unit Price (In Rs.)	GST @ %	Total Price with GST (In Rs.)
01	Oral & Maxillofacial Cross Drive Max Kit.	Specification are attached as annexure - I	01 Set		

**Note: -**

- Vendors are requested to submit detailed technical specification of quoted item along with brochure.
- Quotation without technical specification and brochure will be strictly rejected.
- A compiled statement of technical specifications needs to be submitted along with quotations.

**TERMS OF CONDITIONS:**

- 1) All quotations to be submitted in the name of Director, AIIMS Nagpur only. Quotations not addressed to Director, AIIMS, Nagpur will not be opened and rejected summarily.
- 2) Rates of articles should be quoted per unit item without including GST.
- 3) Taxes leviable should be quoted separately.
- 4) Delivery Period: - Within 30 days from the date of supply order.
- 5) Delivery should be FOR O/o The Director AIIMS Nagpur.
- 6) **Liquidated Damages charges:** - In the event of the Seller's failure supply, Liquidated Damages charges will be deducted to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part their of a week, subject to the maximum value of the Liquidated Damages being not higher that 10% of the value of delayed stores.
- 7) Payment will be released on inspection of supplied material found satisfactory by the concerned department as regard to quality, quantity etc. Broken articles will have to be replaced before payment.
- 8) Bank details such as Account Number, IFSC Code etc. should be furnished so as to facilitate payment online, if any.
- 9) Interested parties should submit their quotations through hard copy in Store, super scribing "QUOTATION FOR SUPPLY OF Oral & Maxillofacial Cross Drive Max Kit for "Department of Dentistry" AT AIIMS, NAGPUR WITH NUMBER OF NIQ" on or before 31/07/2023 before **17.00 hrs.** Which will be opened on 01/08/2023 at **11.00 hrs**
- 10) Rights to accept/reject any quotation rests with the AIIMS Nagpur.

  
Administrative Officer  
AIIMS, Nagpur

Copy for information forwarded to:

- 1) Central Store Depart, AIIMS Nagpur

## "Annexure -I"

### Specification for Oral & Maxillofacial Cross Drive Max Kit

Sr. No.	Items Name	Specification	Qty. Req.
1	Cross Drive Screw Driver Quick Coupling Handle Only	Hard Anodized 100MMx25MM SS Knob	4
2	Cross Drive Screw Driver Shaft Only Non Locking 1.2 MM, 1.5 MM, 2MM, 2.5 MM- each 4	1.5 MM-75MM L , 6MM holding sleeve,2.5MM tip, 2MM- 90MM L , 6MM holding sleeve,3MM tip, 2.5 MM- 90MM L , 6MM holding sleeve,3MM tip	16
3	Self Holding Screw Driver Fibre Handle 1.5 MM, 2MM & 2.5MM Each 2	Shaft Dia: 6mm, Length:6½"	2 Set
4	Ordinary Screw Driver with Fibre Handle 1.5 MM, 2MM & 2.5MM Each 2	Shaft Dia: 4.5mm, Length:6½"	2 Set
5	Trocar Sleeve / Drill Guide	Dia: 2.35mm, Length: 3"	2
6	Depth Gauge 1.5 MM, 2MM & 2.5MM	Tip. Dia : 1mm, 1.5 MM, 2MM Length : 4" Laser marked	2 Set
7	Screws / Plate Holding Forceps	Length: 6", Fenestrated Handle , Fine Tip	2
8	Plate Holding Fork	4mm Round Ball with groove, Length:6"	2
9	Bone Elevator - Hohman Type 6mm	Snake Point Curved 6mm wide , Length:6"	2
10	Bone Elevator - Hohman Type 8mm	Snake Point Curved 8mm wide , Length:6"	2
11	Reduction Bone Holding Forceps	Semi Sharp pointed single jaw with ratchet, Length : 6"	2
12	Plier	Serrated Jaw without gap, Length:5½"	2
13	Modelling Lever	3 pins Dia. 1.5/2/2.5MM, Length:5½"	2
14	Wire Cutter	5" long X 10mm tip hardened, Box joint	2
15	Wire Twisting Forceps	6" long, 8.5mm thick Diamond serration, 3 ratchets 6" long, 8.5mm thin Diamond serration, 3 ratchets	2 2
16	Mini Plate Cutter ( Vertical )	L 145MM 1.5,2.5MM vertical cutting slot	2
17	Arch Bar Cutter	7.5" X 12mm TC tip,0.7mm competible	2
18	Container For Eco Pro/Startup Kit	With Three trays	2
19	SS Drill Bits 1, 1.5 and 2 MM each 5	2.35MM shaft dia,Ultra flute with stopper JPN steel	15
20	Mandibular Reconstruction Plates Bender ( Roller Type )	3 Round Bars Halfmoon Handle , Length:8"	2
21	Multi Plate Cutter ( Circular Type )	L 9" x 1.5,2,2.5,3mm slots with spring	2

8)

### Terms and Conditions

1. All the instruments must be 'CE' certified with European Notified Body.
2. Submit 'CE' certificate with product annex.
3. The bidder must submit certificate of ISO 13485 : 2016
4. It will be at the sole discretion of the authority to select the product after proper approval of samples. The bidder must submit samples of all the quoted items as and when asked.
5. Material- SS-316L - TITA. PLATES GR.II ASTM F67 TITA. SCREWS GR.V ASTM F136
6. Manufacturer must have MD 9 Certificate with Product Permission.
7. Must have Lab testing Report.
8. Manufacturer must have average turnover of 8 Cr for last 3 years.
9. Manufacturer Turnover Certificate should be submitted.
10. Warranty – 01 Year.
11. Compulsory MD-5 Certificate.



**Dr. Ramaraj P.N.**  
MDS. FFDRCSI, MOMS RCPS Glasg  
Reg. No.3773A (KSDC)  
Professor and Head  
Dept. of Dentistry, AIIMS Nagpur