



अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR

Address: Plot No. 2, Sector-20, MIHAN, Nagpur- 441108

E-mail: procurement@aiimsnagpur.edu.in

File No. AIIMS-NAG/PROC/Cent-Store/25-26/34/NIQ/ 13

Dt. 17/10/2025

NOTICE INVITING QUOTATIONS

The Executive Director, All India Institute of Medical Sciences, Nagpur invites quotations from the reputed Procurement of Endoscopy manual cleaning sink with two flushing pumps for Pulmonary Medicine at AIIMS, Nagpur details are given below:

Sr. No.	Item Description	QTY.	Unit Cost (Excluding GST)	Total Cost (Excluding GST)	GST %	Total Cost with GST
1	Endoscopy manual cleaning sink with two flushing pumps	01 Nos				
Total -----						

Note:

- The technical specification of the item is attached as Annexure-I.
- A compiled statement of technical specification against the item needs to be submitted along with quotation.

TERMS OF CONDITIONS:

- 1) All quotations to be submitted in the name of **Executive Director, AIIMS Nagpur** only. Quotations not addressed to Executive Director, AIIMS, Nagpur will not be opened and rejected summarily.
- 2) L1 will be decided based on the cost of equipment only.
- 3) Price of Consumables should be quoted separately.
- 4) **Copy of NIQ duly signed and stamped by the vendor to be submitted along with quotation by accepting all the terms and conditions of NIQ.**
- 5) HSN Code of the product needs to be mentioned.
- 6) Rates of articles should be quoted per unit item without including GST.
- 7) Taxes livable should be quoted separately.
- 8) Delivery Period: - Within 30 days from the date of supply order.
- 9) Delivery should be FOR O/o Central Store, AIIMS, Nagpur.
- 10) **A declaration by vendor is required to be submitted along with quotation station that vendor is not debarred by Department of Commerce or Ministry/ Department concerned. The date of declaration should not be before the date of NIQ and after the last date of submitted of quotation.**
- 11) **Liquidated Damages charges:** - In the event of the Seller's failure supply, Liquidated Damages charges will be deducted to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part them of a week, subject to the maximum value of the Liquidated Damages being not higher that 10% of the value of delayed stores.
- 12) Payment will be released on inspection of supplied material found satisfactory by the concerned department as regard to quality, quantity etc. Broken articles will have to be replaced before payment.
- 13) Bank details such as Account Number, IFSC Code etc. should be furnished so as to facilitate payment on line, if any.
- 14) Interested parties should submit their quotations in **sealed envelope in store office, super scribing "QUOTATION FOR SUPPLY OF 'ENDOSCOPY MANUAL CLEANING SINK WITH TWO FLUSHING PUMPS FOR PULMONARY MEDICINE at AIIMS, NAGPUR WITH NUMBER OF NIQ"** on or before 26/10/2025 before 17.00 hrs which will be opened on 25/10/2025 at 11.00 hrs.
- 15) Rights to accept/reject any quotation rests with the AIIMS Nagpur.



[Handwritten Signature]
17/10/25
Administrative Officer - I
AIIMS, Nagpur

रोहन एल. बागडे / Rohan L. Bagde
प्रशासनिक अधिकारी / Administrative Officer
अ.आ.सं. नागपुर / AIIMS, Nagpur

Sr. No.	Name of Item	Specifications
1	Endoscopy manual cleaning sink with two flushing pumps	<p>BROAD-BASED SPECIFICATIONS FOR BRONCHOSCOPE CLEANING SINK</p> <ol style="list-style-type: none"> 1. The set should have three sinks designed for the soaking, flushing, and rinsing of medical scopes, in compliance with established guidelines. 2. Its dimensions should be 2040mm x 730mm x 1780mm ± 20mm with each sink measuring 530MM x 460MM x 200MM (LWH) 3. The sink body must be constructed using 304 stainless steel that is corrosion resistant and free of deformity at extremes of Indian climatic conditions. 4. Sink should be made of plastic/fiber material that has properties of stain resistance and durability. 5. It should include solid back wall for stability made of 304 Stainless steel. 6. SS tray for holding connectors and accessories. 7. 4 - inch LED lights with 6 Amps switches. 8. Sink should have U-Type faucets. 9. Power Requirement of 230 VAC/50Hz 10. Miniature Circuit Breaker(MCB) of suitable rating and Capacitor Control Breaker (CCB) shall be installed as per load and power factor requirements ensuring safety. 11. The sink unit should come integrate with at least 2 channel flushing pump system. 12. The sink Unit should have scope umbilical cord supports (1 for each sink) 13. PVC ball valve, PVC Fittings, drain hose and water inlet pipe should be provided with the unit. 14. A 3 inch 7- segment digital Clock should be pre-installed with the sink unit. 15. Seller should provide at least two glass lid with holder with the sinks. 16. Shelves with adequate space to be provided under the sink with sliding doors. 17. Must be supplied with all accessories to make the unit functional. 18. Onsite warranty for five years. 19. Service must be provided within 7 days of registration of complaint, failing which penalty will be imposed as per rule. 20. Supplier should arrange for site visit of specialist for demonstration as designated by hospital for technical inspection of product already installed and operational of similar make.

