



# अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR

Address: Plot No. 2, Sector-20, MIHAN, Nagpur- 441108

Email - hospitalstore@aiimsnagpur.edu.in

File No.: ADMIN-18.0/187/2025- O/O AO(ADMIN)-AIIMS-NAGPUR/3315/NIQ/32

Date: 10/11/2025


## NOTICE INVITING QUOTATIONS

The Executive Director, All India Institute of Medical Sciences, Nagpur invites quotations from the reputed supplier on their letter head for supply of **Surgical Dressing Drum** for Department of Trauma & Emergency OT of AIIMS Nagpur mentioning the rate in following format:

Sr. No.	Item Name	QTY	Unit Cost	Total Cost	GST %	Total Cost with GST
1	Surgical Dressing Drum- SS 304 Seamless (Jointless)-Salem Size- 400 mm x 320 mm (15" x 12") Technical specification as per Annexure-1	20				
Total---						

### TERMS OF CONDITIONS:

- 1) All quotations to be submitted in the name of Executive Director, AIIMS Nagpur only. Quotations not addressed to Executive Director, AIIMS, Nagpur will not be opened and rejected summarily.
- 2) **Copy of NIQ duly signed and stamped by the vendor to be submitted along with quotation by accepting all the terms and conditions of NIQ.**
- 3) Rates of articles should be quoted per unit item without including GST.
- 4) **Validity of quotation should be minimum 60 days from the date of quotation.**
- 5) Taxes leviable should be quoted separately.
- 6) Delivery Period: - Within 30 days from the date of supply order.
- 7) Delivery should be FOR Destination i.e. AIIMS, Nagpur
- 8) Liquidated Damages charges: - In the event of the Seller's failure supply, Liquidated Damages charges will be deducted to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part their of a week, subject to the maximum value of the Liquidated Damages being not higher that 10% of the value of delayed stores.
- 9) Payment will be released on inspection of supplied material found satisfactory by the concerned department as regard to quality, quantity etc. Broken articles will have to be replaced before payment.
- 10) Bank details such as Account Number, IFSC Code etc. should be furnished so as to facilitate payment on line, if any.
- 11) **A declaration by vendor is required to be submitted on firm's letter head duly signed & stamped along with quotation stating that "vendor is not debarred by Department of Commerce or Ministry / Department concerned". The date of declaration should not be before the date of NIQ and after the last date of submission of quotation.**
- 12 Interested parties should submit their quotations **within 07 days from the date of issue/uploading of the NIQ.** The quotations should be sent in sealed envelope in "Central Medical Store, 1<sup>st</sup> Floor, IPD Building, AIIMS Nagpur", superscribing on envelop "QUOTATION FOR SUPPLY OF SURGICAL DRESSING DRUM FOR DEPARTMENT TRAUMA & EMERGENCY OT WITH NUMBER OF NIQ" within working hour of Institute. If last date is Sunday or holiday, next working day will be considered as last date of submission.
- 13) Rights to accept/reject any quotation rests with the AIIMS Nagpur.

  
Administrative Officer-I  
रोहन एल. बागडे / Rohan L. Bagde  
पञ्चासन्निक अधिकारी / Administrative Officer  
AIIMS, Nagpur

## Annexure-1

**Dressing Drum – SS 304 Seamless (Jointless) – Salem**  
**Size:** 420 mm x 320 mm [15" x 12"]

### Technical Specifications:

- **Material:** SS 304 Salem (Seamless / Jointless)
- **Body Thickness:** 0.5 mm
- **Flange Thickness:** 0.4 mm
- **Top Lid Thickness:** 0.6 mm
- **Construction:** Heavy gauge with rivets on handle and hinge for extra strength
  - 4 heavy rivets on handle
  - 4 heavy rivets on hinge
- **Embossing:** "SS 304" must be embossed on the drum body
- **Locking System:** Key-type locking system (Button locks not acceptable)
- **Demonstration:** Mandatory at the time of supply

### Compliance & Documentation:

- Material grade certificate for **SS 304 Salem** to be submitted
- Manufacturer authorization letter specifically for **AIIMS Nagpur** (General authorization not acceptable)
- Manufacturer must be **ISO 9001 & CE Certified**