



अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR

Address: Plot No. 02, Sector-20, MIHAN, Nagpur- 441108

Email: - procurement@aiimsnagpur.edu.in

Ph.No.0710-3295590

No. AIIMS/NGP/ Proc/Cent-Store/2025-26/69/Com.Obj/ 4899

Date: 26/02/2026


Invitations for Comments / Objections

Subject: Proposal for Purchase of 'Adult Cardiac Probe/Transducer (Echo probe model rP19x compatible for EDGE II system)' for Department of Anesthesiology for SICU unit at AIIMS Nagpur –Inviting Comments/ Objection thereon.

The Institute is in the process of purchasing 'Adult Cardiac Probe/Transducer (Echo probe model rP19x compatible for EDGE II system)' for Department of Anesthesiology for SICU unit at AIIMS, Nagpur from M/s Vijayalaxmi Enterprises, Nagpur who is authorized supplier of M/s Fujifilm Sonosite India Pvt., Ltd., Gurugram who is subsidiary company of M/s Fujifilm Sonosite Inc., USA. The proprietary certificate and Authorization form submitted is attached & uploaded on website.

The above documents are being uploaded for open information to submit objection / comments, if any from other manufacturer regarding proprietary nature of the above said system and accessories/consumables within **15 days** from the date of issue/uploading of the notification. The comments should be sent to **Central Procurement Cell, Plot No. 02, Sector-20, MIHAN, AIIMS Nagpur-441108** or through Email to procurement@aiimsnagpur.edu.in on or before **13/03/2026** up to 5:00 PM, failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.




Administrative Officer - I
रोहन एल. बागडे / Rohan L. Bagde
प्रशासनिक अधिकारी / Administrative Officer
अ.भा.आ.सं. नागपुर / AIIMS, Nagpur

Enclosure: Related documents enclosed

Proprietary Article Certificate under Rule No. 166 of GFR 2017

Proprietary Article Certificate in the following form is to be provided by the Ministry/Department before procuring the goods from a single source under the provision of sub Rule 166 (i) and 166 (iii) as applicable.

- (i) The indented goods are manufactured by M/s Fujifilm Sonosite India Pvt. Ltd.
- (ii) No other make or model is acceptable for the following reasons:
- a) Compatibility
- b) No other manufacturer provide the ^{above} equipment
- (iii) Concurrence of finance wing to the proposal vide:.....
- (iv) Approval of the competent authority vide:.....

Syeshram
20/8/25

Signature with date and

Designation of Indenting Officer

Dr. Gajanan N. Chavan
26/8/25

Signature, Head of the Department

With Office seal

Date:

डॉ. गजानन न. चव्हाण / Dr. Gajanan N. Chavan
MBBS, MD, FACEE
प्राध्यापक / Professor
निधेतनाविज्ञान एवं क्रिटिकल केअर विभाग
Dept. of Anaesthesiology and Critical Care
एम्स, नागपूर / AIIMS, Nagpur
Rev. No. 90169

ECHO PROBE TECHNICAL SPECIFICATIONS - SICU

Probe / Transducer Name:	rP19x (Adult Cardiac Probe/Transducer.		
Frequency:	5-1MHz Phased Array.		
Application Used For:	Adult Cardiology, Abdominal, Lung, Orbital.		
Application Specific Calculation:	LVO, Automated Cardiac Output Packages & Patient Report Including		
	Ventricular, Aortic & Atrial Measurements; Ejection Fraction, Volume		
	Measurement, Simpson's Rule, Continuity Equation, Pressure Half		
	Time & Cardiac Output; IVC Collapse Ratio, LA/RA Volume, Tapse,		
	PA AT, TV E, A, PHT, TVI, MV Time, Pulm Veins, LV Mass, TDI e', TDI a',		
	HR, dp:dt, Qp/Qs.		
	Ability To view EF & FS Simultaneously		

Sucheta Meshram

19/01/26

Dr. SUCHETA MESHAM
 ADDITIONAL PROFESSOR & SURGICAL ON CHARGE
 DEPT. OF ANAESTHESIOLOGY & CRITICAL CARE
 AIIMS, NAGPUR
 Reg. No: 2003083315

Head Office: FUJIFILM Sonosite India Pvt. Ltd., Knowledge Boulevard, 6th Floor, Tower B, Plot A/8A, Sector-62, Noida, Uttar Pradesh – 201309, India Tel: +91 120 4414100 Email: ffss-india@fujifilm.com

Registered Office: FUJIFILM Sonosite India Pvt. Ltd. Business Centre Office No. 521, Level-V, Caddie Commercial Tower, Hospitality Dist. Aerocity, IGI Airport, New Delhi 110037 IN, CIN: U33112DL2007FTC327990

The Director
AIIMS
Nagpur Maharashtra 441108

Date: 03/09/2025

Sub: Proprietary Article Certificate for Cardiac probes model rP19x for EDGE II system

Dear Sir,

We, FUJIFILM Sonosite India Private Limited, House No.1-A, 2 Khasra No. 2714 2 5, Near Sheetla Gas Godown, Ashok Vihar Phase-3, Gurugram, Haryana-122001 India, a wholly owned subsidiary of M/s FUJIFILM Sonosite Inc., who are established and reputed manufacturer of Compact Color Doppler Ultrasound Systems having factories at 21919 30th Drive SE, Bothell, WA 98021-3904 USA who are established and reputed manufacturer of Compact Color Doppler Ultrasound Systems and phased array transducer model rP19x/5-1MHz is compatible with our Sonosite EDGE II System is our proprietary item and not being manufactured by any other firm.

We declare that no other spares parts other than phased array transducer model rP19x/5-1MHz, would be compatible with Sonosite EDGE II Color Doppler Ultrasound System.

Yours sincerely,

For FUJIFILM Sonosite India Pvt. Ltd.

Hital Singh



Authorized signatory

Head Office: FUJIFILM Sonosite India Pvt. Ltd., Knowledge Boulevard, 6th Floor, Tower B, Plot A/8A, Sector-62, Noida, Uttar Pradesh – 201309, India Tel: +91 120 4414100 Email: ffss-india@fujifilm.com

Registered Office: FUJIFILM Sonosite India Pvt. Ltd. Business Centre Office No. 521, Level-V, Caddie Commercial Tower, Hospitality Dist. Aerocity, IGI Airport, New Delhi 110037 IN, CIN: U33112DL2007FTC327990

The Director
AIIMS
Nagpur Maharashtra 441108

Date: 03/09/2025

Ref: Your requirement of echo probe model rP19x compatible for EDGE II system

Sub: Authorization Letter

Dear Sir,

We, FUJIFILM Sonosite India Private Limited, House No.1-A, 2 Khasra No. 2714 2 5, Near Sheetla Gas Godown, Ashok Vihar Phase-3, Gurugram, Haryana-122001 India, a wholly owned subsidiary of M/s FUJIFILM Sonosite Inc., who are established and reputed manufacturer of Compact Color Doppler Ultrasound Systems having factories at 21919 30th Drive SE, Bothell, WA 98021-3904 USA do hereby authorize M/s. Vijayalakshmi Enterprises, Plot No. 68, Rajabaksha, Near Hanuman Temple, Medical Square, Nagpur, Maharashtra - 440003 to submit an offer for phased array transducer model rP19x/5-1MHz for EDGE II system and subsequently negotiate, sign the Supply order with you against your particular this requirement.

No company or firm or individual other M/s. Vijayalakshmi Enterprises, Plot No. 68, Rajabaksha, Near Hanuman Temple, Medical Square, Nagpur, Maharashtra - 440003 are authorized to bid, negotiate and conclude the contract in regard to this business.

Yours faithfully,

For **FUJIFILM Sonosite India Pvt. Ltd**

Hital Singh

Authorized Signatory

