



अखिल भारतीय आयुर्विज्ञान संस्थान, नागपुर

ALL INDIA INSTITUTE OF MEDICAL SCIENCES, NAGPUR

Address: Plot No. 2, Sector-20, MIHAN, Nagpur- 441108

Email - hospitalstore@aiimsnagpur.edu.in

File No. ADMIN-18.0/19/2026-O/O AO(ADMIN)-AIIMS-NAGPUR /3575/NIQ/ 60

Date: 09/03/2026

NOTICE INVITING QUOTATIONS

The Executive Director, All India Institute of Medical Sciences, Nagpur invites quotations from the reputed supplier on their letter head for **Supply of EMG Needle Electrodes for Department of Physiology** at AIIMS Nagpur mentioning the rate in following format:

Sr. No	Item Description	QTY	Unit Cost	Total Cost	GST %	Total Cost with GST
1	EMG Needle Electrodes (Length:2 Inch) (Technical Specification as per Annexure-I)	100				
2	EMG Needle Electrodes (Length:1.5 Inch) (Technical Specification as per Annexure-I)	50				
Total--						

Note- L1 will be decided item wise including GST

TERMS OF CONDITIONS:

- 1) All quotations to be submitted in the name of Executive Director, AIIMS Nagpur only. Quotations not addressed to Executive Director, AIIMS, Nagpur will not be opened and rejected summarily.
- 2) Copy of NIQ duly signed and stamped by the vendor to be submitted along with quotation by accepting all the terms and conditions of NIQ.
- 3) Rates of articles should be quoted per unit item without including GST.
- 4) Validity of quotation should be minimum 60 days from the date of quotation.
- 5) Taxes leviable should be quoted separately.
- 6) Delivery Period: - Within 30 days from the date of supply order.
- 7) Delivery should be FOR Destination i.e. AIIMS, Nagpur
- 8) Liquidated Damages charges: - In the event of the Seller's failure supply, Liquidated Damages charges will be deducted to the sum of 0.5% of the contract price of the delayed/undelivered, stores/services mentioned above for every week of delay or part their of a week, subject to the maximum value of the Liquidated Damages being not higher than 10% of the value of delayed stores.
- 9) Payment will be released on inspection of supplied material found satisfactory by the concerned department as regard to quality, quantity etc. Broken articles will have to be replaced before payment.
- 10) Bank details such as Account Number, IFSC Code etc. should be furnished so as to facilitate payment online, if any.
- 11) A declaration by vendor is required to be submitted on firm's letter head duly signed & stamped along with quotation stating that "vendor is not debarred by Department of Commerce or Ministry / Department concerned". The date of declaration should not be before the date of NIQ and after the last date of submission of quotation.
- 12) Interested parties should submit their quotations within 07 days from the date of issue of the NIQ. The quotations should be sent in sealed envelope in "Central Medical Store, 1st Floor, IPD Building, AIIMS Nagpur", superscribing on envelop "QUOTATION FOR SUPPLY OF EMG NEEDLE ELECTRODES FOR DEPARTMENT OF PHYSIOLOGY OT WITH NUMBER OF NIQ" within working hour of Institute. If last date is Sunday or holiday, next working day will be considered as last date of submission.
- 13) Rights to accept/reject any quotation rests with the AIIMS Nagpur.


Administrative Officer-I
रोहन एल. बागडे
प्रशासनिक अधिकारी / Administrative Officer
अ.मा.आ.सं. नागपुर / AIIMS, Nagpur

DETAILED TECHNICAL SPECIFICATIONS**1. EMG Needle Electrodes (Length: 2 Inch)**

- **Cannula** : Stainless Steel Canula
- **Coating** : Silicone
- **Core** : Tungsten
- **Needle Recording Area** : 0.07 mm²
- **Gauze and Length** : 26 Gauze and 2 Inch/50mm
- **Needle Tip** : Three Bevel Tip
- **Insulator** : Polyesterimide
- **Hub** : Brass, Copper, or Nickel Plating
- **Hub Cover** : Must have Color code, Security Clip and Recording Area Indicator
- **Protective Tube** : Polyethylene
- **Shelf-Life** : > 18 Months during delivery date.

2. EMG Needle Electrodes (Length: 1.5 Inch)

- **Cannula** : Stainless Steel Canula
- **Coating** : Silicone
- **Core** : Tungsten
- **Needle Recording Area** : 0.07 mm²
- **Gauze and Length** : 26 Gauze and 1.5 Inch/37mm
- **Needle Tip** : Three Bevel Tip
- **Insulator** : Polyesterimide
- **Hub** : Brass, Copper, or Nickel Plating
- **Hub Cover** : Must have Color code, Security Clip and Recording Area Indicator
- **Protective Tube** : Polyethylene
- **Shelf-Life** : > 18 Months during delivery date.

Remarks: A sample piece will be required during the technical evaluation.